Defender™ Instructions for Use

Product Number: FS3


Assembly

The Defender[™] foot module is pre-assembled consisting of a graphite component (keel), Spectra[™] sock, and foot shell. Foam *stiffening bumpers* for increasing heel stiffness are provided. After dynamic alignment, torque pyramid adjustment screws to the manufacturer's specifications. Secure pyramid adjustment screws with a thread locking adhesive (i.e., Loctite 242).

Bench Alignment

Prior to donning the prosthesis:

- · Plantarflex/Dorsiflexion foot to match the shoe heel height.
- Adduct/Abduct socket to provide appropriate frontal plane angle.
- Flex/Extend socket to provide appropriate sagittal plane angle.
- Move the socket linearly to ensure the weight line falls along the anterior edge of the pylon (see illustration).



Dynamic Alignment

During *loading response*, the heel lever stores energy and releases it during *midstance*. This action provides momentum for the keel to store energy and release it during *terminal stance*. To optimize the heel to toe rollover motion, adjust the following variables:

- Anterior/posterior foot placement
- Dorsiflexion/plantarflexion
- · Heel stiffness

Troubleshooting

Heel too soft

- SymptomsFoot flat occurs too rapidlyToe feels excessively stiffKnee hyperextension
- Solutions
- Shift socket anteriorly in relation to the foot
- · Attach foam stiffening bumpers

Heel too hard

Symptoms

- Rapid knee flexion, instabilityHeel to toe progression to rapidLack of energy return sensation
- Solutions
- Shift socket posteriorly in relation to the footVerify appropriate foot module category
- Verify appropriate root module categ

Foot module too stiff

Symptoms

- Flat spot in rollover motion at slow cadences Solutions
- Consider a lower category foot module

Foot module too soft

Symptoms

- · Clicking noise at initial contact
- Excessive toe deflection with high impact activity Solutions

Consider a higher category foot module

Stiffening Bumpers

Foam stiffening bumpers are included to adjust the heel stiffness during loading response. The bumpers may be temporarily attached with tape between the heel lever and keel. If the bumpers provide excessive heel stiffness, exclude or trim as appropriate. For permanent placement, adhere bumpers using Barge or similar contact cement.

Spectra™ Sock

A Spectra[™] sock is provided to protect the foot shell and minimize noise. Spectra[™] socks must be replaced at intervals appropriate to the user's activity level. Failure to inspect and replace the Spectra[™] socks may prematurely wear the foot module, and will void the warranty.

Foot Shell

When removing or installing the foot shell, use the Foot Shell Removal Tool (ACC-00-10200-00) to prevent damage to the foot module.

Defender™ System

Minimum clearance: 94mm-118mm Maximum user weight: 45 kg (100 lbs)

Available sizes: 16cm-21cm

Heel height: 3/8"

Warranty: Graphite components/pyramid connector (18 months)

Foot shell (12 months)

Maintenance

The foot module requires periodic maintenance.

- Inspect the foot module every six months. If the user is more active, more frequent inspection may necessary. Service as necessary. Replace Spectra™ sock and/or foot shell if worn to prevent damage to the graphite components.
- The foot module may be cleaned and/or disinfected with soap and warm water. If the foot is exposed to water, remove the foot shell to clean and disinfect its interior.

Warnings

Failure to adhere to the guidelines of the *Instructions for Use* will void the warranty.

- Never use the foot module without a *foot shell*. Failure to comply may cause premature wear, loss of function, and/or product failure.
- Always use the foot module with a sock and shoe. Failure to comply may cause premature wear, loss of function, and/or product failure.
- Do not allow aggregates such as sand to remain in the foot shell. Upon exposure to aggregates, immediately disassemble foot module and rinse with water. The abrasive properties of aggregates will quickly wear the graphite components of the foot module.
- Freedom Innovations foot modules are manufactured to fit industry standard pyramids and receivers. It is the prosthetist's responsibility to select and/or fabricate properly fitting attachment components.
- Never attempt to loosen the bolt affixing the pyramid connector.
- Discontinue use and consult your prosthetist if any part of the prosthesis starts to make noise.
- Inform your prosthetist if you lose or gain a significant amount of weight.
- Freedom Innovations foot products are manufactured and tested for a particular weight and activity impact level. Use by another user for whom it was not originally manufactured may cause injury and shall void any written or implied warranty.





Advena Ltd. Pure Offices, Plato Close, Warwick, CV34 6WE UK